



**CENTRAL INSTITUTE OF HIGHER TIBETAN STUDIES, SARNATH VARANASI**  
( Deemed University )

~Enquiry details Register~

Enquiry No. : CIHTS/PPC/18-19-11  
Enquiry Notice

Dated: 12-03-2019

Sub. : Quotation for below described Items.


Dear Sir,

You are invited to offer your lowest quotation for the following equipment/materials as per the terms and conditions given below and overleaf:

Sl. No.	Name of Items	Qty	UOM
1.	<b>SEMI AUTOMATIC BIOCHEMISTRY ANALYZER WITH FOLLOWING SPACIFICATION:</b> 1.It Should be microprocessor controlled programmable, semi autoanalyser to perform routine biochemistry tests, enzyme immunoassays. 2. The unit should be able to offer 10 modes such as 1-Point Linear (End-Point), 2-Point Linear (Fixed Time), Rate A Linear (Kinetic), 1-Point Non-linear (End-Point), 2-Point Non-linear (Fixed Time), Rate A Non-linear (Kinetic), 1-Point Sample Blank Linear (Differential), 1-Point Sample Blank Non-linear (Differential), Absorbance & Coagulation 3. It should have facility to use both 6mm glass cuvettes and 10mm plastic cuvettes additionally in place of flow cell & there should be a READ button to take readings from 6mm glass cuvettes and 10mm plastic cuvettes 4. It should have 8 narrow band static interference filters with wavelength selectable from 340 - 670 nm. Without any use of a moving filter wheel. 5. The instrument should also be capable of doing coagulation assays with programmable ISI value to print results in INR. 6. It should have facility to store minimum 1000 test results in the memory. Further it should possible to recall the stored results with Patient ID or with Date or with Date & Patient ID. 7.The manufacturer / supplier should have a full-fledged service force and the manufacturer should be able to supply kits locally against orders. 8. The analyzer when is switched on it should display which filter of photometer is OK or Not OK . 9. It should have Self Diagnosis software for easy maintenance of analyzer. 10. Analyzer should provide help key facility for easy performance of a test.	01	Nos.
2	<b>Centrifuge with 8 tube facility</b>	01	Nos

**TERMS AND CONDITIONS**

- (a) The firms or supplier should submit their proper authorization of company.  
(b) Price quoted for the material/equipment shall be firm and valid for 60 days from the date of opening.  
(c) The material/equipment offered shall confirm to relevant Indian/International standards.
- Delivery shall be essentially:
  - At CIHTS Stores for local offers.

Sd/-   
Authorized Signatory

(b) For Varanasi Railway Station for outstation offers.(for outstation offers approximate freight by Road/Railways by passenger/Goods train shall be indicated if the offer is not F.O.R. varanasi.)

3. Rate offered must be for metric system of unit lengths or volume or weight.
4. The Quotation shall be complete in all respects and the firms shall give the brand name offered by item in the quotations. Wherever applicable, technical literature may please be enclosed along with tender.
5. CIHTS reserves the right to increase the quantity mentioned in this enquiry and CIHTS reserves the right to split this tender and place order on one or more tenderers and right to reject partly or completely and tender without assigning any reasons thereof.

**Note:** Last date of receiving quotation is 25-03-2019 upto 3.30 p.m. The tender will be opened at 3.30 p.m. on the next working day.. Please submit your quotation in duplicate in sealed cover super scribing the Tender Enquiry number and date.

---

